# Monarch fixed ceiling hoist

	$\sim$		~=	

Product weight	Manual model: 7.5 kg (17 lb); motorized model: 8.5 kg (19 lb)
Product life expectancy	10 years with only 1 battery replacement
Safe working load (SWL)	272 kg (600 lb), 200 kg (440 lb), or 130 kg (286 lb)
Lifting speed	5.5 cm/sec (2.2 in/sec) at 0 kg; 5 cm/sec (2 in/sec) at 130 kg (286 lb) 4.5 cm/sec (1.8 in/sec) at 200 kg (440 lb); 4 cm/sec (1.6 in/sec) at 272 kg (600 lb)
Water ingress protection rating of lift	IP21
Noise level	Maximum 54 dBA
Medical electrical equipment class	Class 1
Protection class	Type BF
Warranty	2 years

#### **BATTERY**

Lithium ion, 25.2V 2500 mAh
Approximately 80 cycles of 61 cm (24 in) at 91 kg (200 lb)
Approximately 70 cycles of 61 cm (24 in) at 130 kg (286 lb)
Approximately 45 cycles of 61 cm (24 in) at 200 kg (440 lb)
Approximately 30 cycles of 61 cm (24 in) at 272 kg (600 lb)
Full capacity in approximately 2 hours

## **CHARGER**

Charger input	100-240 VAC, 50-60Hz
Charger output	29.4 VDC, max 1A

## **ENVIRONMENTAL CONDITIONS**

Operating temperature	5° C to 40° C (41° F to 104° F)
Operating humidity	15 to 93%, non-condensing
Storage temperature	-25° C to 70° C (-13° F to 158° F)
Storage humidity	0 to 93%, non-condensing
Atmospheric pressure	70 kPa to 106 kPA

## **CONTROLS**

Lift controls	UP and DOWN buttons to raise and lower the lift	
Hand control	UP, DOWN, LEFT and RIGHT buttons to raise and lower the lift,	
mand control	and move along the track	
Hand control operating force	4–4.5N	
Water ingress protection rating of hand control	IP67	
On/Off power switch	Used to shut off power to the lift and as an emergency stop feature	

## **CERTIFICATION**

• ISO 10535:2006 (patient lift)	• IEC 60601-1:2005 A1:2012 (medical electrical equipment)
<ul> <li>CAN/CSA Z10535.1:15 (patient lift)</li> </ul>	• IEC 60601-1-11:2015 (homecare)

## **COMPLIANCE**

CE Marking per 93/42/EEC (medical device)	• 2011/65/EU (ROHS – 100% of components)
<ul> <li>2006/42/EC (machinery directive)</li> </ul>	• 2002/96/EC (WEEE)

**WARNING:** This equipment is not suitable in the presence of flammable anesthetic mixtures with air or oxygen, or with nitrous oxide.